

CERTIFICATE OF ANALYSIS

For

Ultra Ninhydrin Solution NC

Part Number: 80-6003-08

Batch (Control) Number: P27772NC

Quantity: 250 mL

Expiry Date: 13 October 2024

The above product gave the following results on analysis:

Baseline Drift:

0.008 $\Delta mV/\Delta t$ at 440nm, 0.030 $\Delta mV/\Delta t$ at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

Colorimetric Test:

Absorption @ 485nm = 0.038 A

Absorption @ 600nm = 0.002 A

Absorption Specification <0.10 at 485nm, <0.05 at 600nm

Chemical QC Signature:

Date of Issue: 19 October 2021

David Thomas

| | | | |
|---|---|--------------------|------------------------|
| Postal Address | Telephone | Telefax | Email: |
| Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England. | +44 (0)1223 423723 | +44 (0)1223 420164 | support@biochrom.co.uk |
| | Web Site: http://www.biochrom.co.uk | | |

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.