

CERTIFICATE OF ANALYSIS

For

Sodium Accelerated Buffer 4

Part Number: 80-2115-24

Batch (Control) Number: U27890

Quantity: 2 Litres

Expiry Date: 11 June 2027

The above product gave the following results on analysis:

pH = 9.219 @ 20.0°C pH Specification = 9.21 - 9.25 @ 20.0°C

Conductivity = $48.5 \text{ mScm-1} @ 20.0^{\circ}\text{C}$ Conductivity Specification = $47.0 - 51.0 \text{ mScm-1} @ 20.0^{\circ}\text{C}$

Chemical QC Signature: Date of Issue: 21 June 2024

Neil Booth

Postal Address Telephone

+44 (0)1223 423723

Telefax +44 (0)1223 420164 Email:

support@biochrom.co.uk

Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach

Cambridge

Biochrom Ltd

CB25 9PE, England. Web

Web Site: http://www.biochrom.co.uk

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.

Certificate of Analysis v1.1 - 29 September 2020.