

CERTIFICATE OF ANALYSIS

For

Ultra Ninhydrin Solution NC

Part Number: 80-6003-08

Batch (Control) Number: U27903NC

Quantity: 250 mL

Expiry Date: 30 October 2027

The above product gave the following results on analysis:

Baseline Drift:

-0.06 $\Delta mV/\Delta t$ at 440nm, 0.03 $\Delta mV/\Delta t$ at 570nm

Drift Specification <0.09 at 440nm, <0.06 at 570nm

Colorimetric Test:

Absorption @ 485nm = 0.055 A

Absorption @ 600nm = 0.009 A

Absorption Specification <0.10 at 485nm, <0.05 at 600nm

Chemical QC Signature:

Date of Issue: 04 November 2024

David Thomas

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
Web Site: http://www.biochrom.co.uk			

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.