

CERTIFICATE OF ANALYSIS

For

Lithium Buffer 1 NC

Part Number: 80-6003-01

Batch (Control) Number: U27908NC

Quantity: 2 Litres

Expiry Date: 06 December 2027

The above product gave the following results on analysis:

pH = 2.806 @ 20.0°C

pH Specification = 2.78 - 2.82 @ 20.0°C

Conductivity = 14.66 mScm-1 @ 20.0°C

Conductivity Specification = 14.4 - 16.0 mScm-1 @ 20.0°C

Chemical QC Signature:

Date of Issue: 11 December 2024

David Thomas

Postal Address

Telephone

Telefax

Email:

Biochrom Ltd
Unit 7, Enterprise Zone
3970 Cambridge Research Park
Beach Drive, Waterbeach
Cambridge
CB25 9PE, England.

+44 (0)1223 423723

+44 (0)1223 420164

support@biochrom.co.uk

Web Site: <http://www.biochrom.co.uk>

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.