

CERTIFICATE OF ANALYSIS

For

Ultrosolve Plus

Part Number: 80-2117-65

Batch (Control) Number: W27918

Quantity: 1.75 Litres

Expiry Date: 02 April 2028

The above product gave the following results on analysis:

pH = 5.632 @ 20.0°C

pH Specification = 5.55 - 5.75 @ 20.0°C

Conductivity = 19.1 mScm⁻¹ @ 20.0°C

Conductivity Specification = 17.5 - 23.5 mScm⁻¹ @ 20.0°C

Chemical QC Signature:

Date of Issue: 10 April 2025

Neil Booth

Postal Address	Telephone	Telefax	Email:
Biochrom Ltd Unit 7, Enterprise Zone 3970 Cambridge Research Park Beach Drive, Waterbeach Cambridge CB25 9PE, England.	+44 (0)1223 423723	+44 (0)1223 420164	support@biochrom.co.uk
Web Site: http://www.biochrom.co.uk			

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.