

CERTIFICATE OF ANALYSIS

For

Sodium Accelerated Buffer 1

Part Number: 80-2115-21

Batch (Control) Number: W27923

Quantity: 2 Litres

Expiry Date: 08 July 2028

The above product gave the following results on analysis:

pH = 3.243 @ 20.0°C

pH Specification = 3.23 - 3.27 @ 20.0°C

Conductivity = 9.60 mScm-1 @ 20.0°C

Conductivity Specification = 9.0 - 11.0 mScm-1 @ 20.0°C

Chemical QC Signature:

Date of Issue: 09 July 2025

David Thomas

Postal Address

Telephone

Telefax

Email:

Biochrom Ltd
Unit 7, Enterprise Zone
3970 Cambridge Research Park
Beach Drive, Waterbeach
Cambridge
CB25 9PE, England.

+44 (0)1223 423723

+44 (0)1223 420164

support@biochrom.co.uk

Web Site: <http://www.biochrom.co.uk>

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.