

# CERTIFICATE OF ANALYSIS

For

## Ultrosolve Plus

Part Number: 80-2117-65

Batch (Control) Number: W27925

Quantity: 1.75 Litres

Expiry Date: 23 July 2028

The above product gave the following results on analysis:

pH = 5.649 @ 20.0°C

pH Specification = 5.55 - 5.75 @ 20.0°C

Conductivity = 19.3 mScm<sup>-1</sup> @ 20.0°C

Conductivity Specification = 17.5 - 23.5 mScm<sup>-1</sup> @ 20.0°C

Chemical QC Signature:

Date of Issue: 28 July 2025

Neil Booth

|                                                                                                                                       |                    |                    |                        |
|---------------------------------------------------------------------------------------------------------------------------------------|--------------------|--------------------|------------------------|
| Postal Address                                                                                                                        | Telephone          | Telefax            | Email:                 |
| Biochrom Ltd<br>Unit 7, Enterprise Zone<br>3970 Cambridge Research Park<br>Beach Drive, Waterbeach<br>Cambridge<br>CB25 9PE, England. | +44 (0)1223 423723 | +44 (0)1223 420164 | support@biochrom.co.uk |
| Web Site: <a href="http://www.biochrom.co.uk">http://www.biochrom.co.uk</a>                                                           |                    |                    |                        |

Registered in England No: 3526954 Registered Office: Unit 7, Enterprise Zone, 3970 Cambridge Research Park, Beach Drive, Waterbeach, Cambridge, CB25 9PE England.